

Product Information

Product Name:	PlexPCR® Colour Compensation
Catalogue Number:	90001
Product Description:	The PlexPCR® Colour Compensation kit is for the generation of a PlexPCR® CC file on the LC480 I, LC480 II and z 480 instruments.
Lot Number:	25010047
Manufacture Date:	2025-07-23
Expiry Date:	2026-07-19
Storage Conditions:	-15°C to -25°C

Test method:

The **PlexPCR®** Colour Compensation kit was tested by creating a CC file and comparing that file to CC file generated by a previously released lot of **PlexPCR®** Colour Compensation.

Results:

Test Samples	Specification	Result
CC file	Comparable with previously released PlexPCR® Colour Compensation lot CC file.	PASS

The Finished products were tested according to standard Quality Control methods. The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

Quality Assurance Signature and Date:

QA Approval	 <small>Lexin Zhang (Jul 25, 2025 09:45:41 GMT+10)</small>
-------------	--