

## Product Information

Product Name:	<b>ResistancePlus<sup>®</sup> MG<sub>(550)</sub></b>
Catalogue Number:	<b>2000201</b>
Product Description:	Detection of <i>Mycoplasma genitalium</i> and 4 mutations in 23S rRNA on the ABI 7500 Fast. This product is approved for <i>in vitro</i> diagnostic use.
Lot Number:	26010029
Manufacture Date:	2026-04-24
Expiry Date:	2028-03-11
Storage Conditions:	<b>-25°C to -15°C</b>

## Test method:

**ResistancePlus<sup>®</sup> MG<sub>(550)</sub>** was tested with reference materials containing *M. genitalium* gene and 23S rRNA mutation sequences using the ABI 7500 Fast instrument with thermal cycling conditions stated in the IFU. The amplifications data was analysed using the **ResistancePlus<sup>®</sup> MG (7500)** software for the interpretation of *M. genitalium* presence and 23S rRNA mutation status.

## Results:

Test Samples	Specification	Result
	Software Interpretation	
<i>M. genitalium</i> , 23S rRNA wild type - 20 copies	<b>Detected:</b> <i>M. genitalium</i>	<b>PASS</b>
<i>M. genitalium</i> , 23S rRNA A2059G - 20 copies	<b>Detected:</b> <i>M. genitalium</i> , 23S rRNA mutation	<b>PASS</b>
<i>M. genitalium</i> , 23S rRNA A2058G - 24 copies	<b>Detected:</b> <i>M. genitalium</i> , 23S rRNA mutation	<b>PASS</b>
<i>M. genitalium</i> , 23S rRNA A2058C - 20 copies	<b>Detected:</b> <i>M. genitalium</i> , 23S rRNA mutation	<b>PASS</b>
<i>M. genitalium</i> , 23S rRNA A2058T - 20 copies	<b>Detected:</b> <i>M. genitalium</i> , 23S rRNA mutation	<b>PASS</b>
Internal Control	<b>Control Valid</b>	<b>PASS</b>
No Template Control	<b>Control Valid</b>	<b>PASS</b>

The Finished products were tested according to standard Quality Control methods. The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

## Quality Assurance Signature and Date:

QA Approval	<u>Lexin Zhang</u> <small>Lexin Zhang (Apr 29, 2026 16:19:08 GMT+10)</small>
-------------	---