

Product Information

Product Name:	PlexPCR[®] RespiVirus Positive Control
Catalogue Number:	95002
Product Description:	The PlexPCR[®] RespiVirus Positive Control kit contains amplification controls. The positive controls contain synthetic DNA for Influenza A (FluA), Influenza B (FluB), Rhinoviruses (RhV), and Respiratory Syncytial Viruses A and B (RSV), Human metapneumovirus (HMPV), Adenoviruses B and C (AdV) and Human parainfluenza viruses 1, 2, 3 and 4 (HPIV 1-4) targets.
Lot Number:	25020048
Manufacture Date:	2025-06-27
Expiry Date:	2027-02-02
Storage Conditions:	-15°C to -25°C

Test method:

PlexPCR[®] RespiVirus Positive Control Kit components were tested with the RV1 and RV2 mixes from the **PlexPCR[®]** RespiVirus₍₆₁₀₎ product and using the LightCycler 480II instrument with thermal cycling conditions stated in the IFU. The amplification data was analysed using the **PlexPCR[®]** RespiVirus (LC480) software for the positive identification of the positive controls.

Results:

Test Sample	Reaction Mix	Specification	Result
RV PC1	RV1	Positive Control 1 - Positive	PASS
	RV2	Positive Control 1 - Positive	PASS
RV PC2	RV1	Positive Control 2 – Positive	PASS
	RV2	Positive Control 2 – Positive	PASS
RV PC3	RV2	Positive Control 3 - Positive	PASS
RV PC4	RV2	Positive Control 4 - Positive	PASS

The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

Quality Assurance Signature and Date:

QA Approval	 Lexin Zhang (Jul 1, 2025 17:09 GMT+10)
-------------	--