

Product Information

Product Name:	PlexPCR® HSV-1&2, VZV
Catalogue Number:	11101L-01
Product Description:	A qualitative qPCR assay for the detection of Herpes simplex virus 1 (HSV-1), Herpes simplex virus 2 (HSV-2), Varicella zoster virus (VZV) on the LightCycler 480II. This product is for in vitro diagnostic use.
Lot Number:	26020025
Manufacture Date:	2026-04-14
Expiry Date:	2028-04-13
Storage Conditions:	-15°C to -25°C

Test method:

PlexPCR® HSV-1&2, VZV was tested with reference materials containing HSV 1 gpd gene; HSV 2 gpd gene and the VZV pol gene using the LightCycler 480II with thermal cycling conditions stated in the IFU.

Results:

Test Samples	Specification	Result
HSV 1_gpd	Detection down to 20 copies	PASS
HSV 2_gpd	Detection down to 16 copies	PASS
VZV_pol	Detection down to 10 copies	PASS
Internal Control	Average Cq within specified range	PASS
No Template Control	No observable amplification	PASS

The Finished products were tested according to standard Quality Control methods. The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

Quality Assurance Signature and Date:

QA Approval	 <small>Lexin Zhang (Apr 20, 2026 09:21:02 GMT+10)</small>
-------------	--