

Product Information

Product Name:	Positive Control HSV/VZV/TP
Catalogue Number:	95007
Product Description:	The Positive Control HSV/VZV/TP swab is formulated with inactivated herpes simplex virus 1 (HSV-1), herpes simplex virus 2 (HSV-2), varicella zoster virus (VZV) and cells containing <i>Treponema pallidum</i> (etiologic agent of Syphilis) genome. The Positive Control HSV/VZV/TP kit can be used as a full process control with nucleic acid detection assays, for use from sample extraction and purification to amplification and detection.
Lot Number:	03200003
Manufacture Date:	11-Jun-2024
Expiry Date:	11-Jun-2026
Storage Conditions:	2-30°C

Quality Control Information

Visual Inspection:

Number of swabs inspected:	Seal pouch inspection	Component controls	Label printing verification
20	PASS	PASS	PASS

Number of swabs tested:	HSV 1	HSV 2	VZV	TP
03	+	+	+	+

Test Results:

Target	Result
Herpes Simplex Virus (HSV) 1	Within specified Cq range
Herpes Simplex Virus (HSV) 2	Within specified Cq range
Varicella Zoster Virus (VZV)	Within specified Cq range
<i>Treponema pallidum</i> (TP) genome	Within specified Cq range

The finished products were tested according to standard Quality Control methods. The results were verified against the design specifications, acceptance criteria and reviewed and approved by the Quality Department.

Quality Assurance Signature and Date:

QA Approval	 <small>Regine Citra (Sep 10, 2024 06:16 GMT+7)</small>
-------------	---